

## **SILICONE VACUUM TUBING**

**\*\*\*CONTENTS NOT STERILE\*\*\***

Package Contains One: #SH-323, 3/8" X 8' (16mm O.D. X 8mm I.D. X 244cm)

**\*\*\*CAUTION\*\*\*** Federal (United States) Law restricts this device to sale by or on order of a physician.

- This tubing is non-conductive and should not be used in the presence of flammable anesthetics.
- This tubing is manufactured as a latex free product.

Each institution should establish the efficiency of their own sterilization procedures. The following techniques have been proved effective and are provided as a guide.

Remove the product from its package in a clean environment. Wash thoroughly in a hot water-soap solution to remove surface contamination. Do not use synthetic detergents or oil-based soaps, as they may be absorbed and leach out during use. Rinse thoroughly in hot water. Wrap in a lint-free cloth or sterilizing paper and place in a clean open tray. Sterilize at 250° F. and at 15psi for 30 minutes.

**\*\*\*NOTE\*\*\*** Properly cleaned, the tubing is suitable for reuse.

Rev 2 - 8/07



**Grams Medical**

2443 Norse Ave – Costa Mesa, CA 92627  
(949) 548-7337 – FAX (949) 548-1056

[sales@gramsmedical.com](mailto:sales@gramsmedical.com) – [www.gramsmedical.com](http://www.gramsmedical.com)

## **SILICONE VACUUM TUBING**

**\*\*\*CONTENTS NOT STERILE\*\*\***

Package Contains One: #SH-3/8, 3/8" X \_\_\_\_\_ ' (16mm O.D. X 8mm I.D. X \_\_\_\_\_ cm)

**\*\*\*CAUTION\*\*\*** Federal (United States) Law restricts this device to sale by or on order of a physician.

- This tubing is non-conductive and should not be used in the presence of flammable anesthetics.
- This tubing is manufactured as a latex free product.

Each institution should establish the efficiency of their own sterilization procedures. The following techniques have been proved effective and are provided as a guide.

Remove the product from its package in a clean environment. Wash thoroughly in a hot water-soap solution to remove surface contamination. Do not use synthetic detergents or oil-based soaps, as they may be absorbed and leach out during use. Rinse thoroughly in hot water. Wrap in a lint-free cloth or sterilizing paper and place in a clean open tray. Sterilize at 250° F. and at 15psi for 30 minutes.

**\*\*\*NOTE\*\*\*** Properly cleaned, the tubing is suitable for reuse.

Rev 2 - 8/07



**Grams Medical**

2443 Norse Ave – Costa Mesa, CA 92627  
(949) 548-7337 – FAX (949) 548-1056

[sales@gramsmedical.com](mailto:sales@gramsmedical.com) – [www.gramsmedical.com](http://www.gramsmedical.com)